

Audiometric Test Room Evaluation

Company: OHM Demo Company Location: Rock Hill, SC

Sound Measurement Equipment: QUEST NLM MOD. 215 Serial Number: M2030053

Date: 12/31/04 Time: 7:00 AM By: OHM User

Test Room Information:

Make: ECKEL
 Model: AB-2000T
 Serial Number: 206128
 Exam Room: 1
 Physical Condition of Test Room:

Door Gasket: OK
 Door Closure: OK
 Ventilation: OK
 Jacks/Plugs: OK
 Earphone Hook: OK
 Lighting: OK
 Seating: OK
 Other: OK

Background

Noise

Measurements

Octave Bands - Center Frequency in Hz

	500	1000	2000	4000	8000	A-Scale
Maximum Allowable Sound Levels in dB SPL (OSHA 29 CFR 1910.95)	40	40	47	57	67	N/A
Outside Test Room	40	40	46	57	60	
Inside Test Room	<30	<30	<30	<30	35	
Other						

Comments/Recommendations: